


National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Page 1 of 1

CALIBRATION LABORATORIES

NVLAP LAB CODE 200459-0

TECHNICAL CALIBRATION LABS, INC.

581 Route 17M
Monroe, NY 10950
Mr. John Tasca
Phone: 845-783-4235 Fax: 845-783-3573
E-Mail: techni@frontiernet.net
URL: <http://www.technicalibration.com>

OPTICAL RADIATION

NVLAP Code: 20/O02

Color Temperature

Range in °Kelvin

Best Uncertainty (±) in °Kelvin^{note 1}

2300

42

2856

54

1. Represents an expanded uncertainty using a coverage factor, $k=2$.

December 31, 2004

Effective through

A handwritten signature in black ink, appearing to read "William R. Muhl".

For the National Institute of Standards and Technology